

Certificate of Analysis

Product: 0.9% Sodium Chloride Injection, USP (AUTO-C)

Lot #: Y303859

Code: 2B1323N

Manufacture/

Sterilization Date: 04/13/2019

Expiry: 07/2020

Chemical Testing per Specification: **03-15-19-061**

Issue Date: **11/13/2018**

TEST	LIMIT	RESULT
NaCl (g/L)	8.55 - 9.45 g/L	8.99 g/L
Sodium ID	Positive	Positive
pH at 25 deg. C	4.5 - 7.0	5.6
Particle Analysis	NMT 25 ≥ 10 um/ml	NMT 25
Particle Analysis	NMT 3 ≥ 25 um/ml	NMT 3
Sterility	Pass Parametric Release	Pass

Endotoxin Testing per Product Control Specification

RESULT	LIMIT	DISPOSITION	DATE PASSED
< 0.25 EU/ml	< 0.25 EU/ml	Pass	4/16/2019

This is to certify that this product was manufactured according to current GMP and fulfills the requirements of the Master Production Document.

Date Batch Released	Parametric Release Date	Quantity Released
4/16/2019	4/16/2019	NA

	Print Name	Signature/Date
Prepared by:	Diana Bartlett	Diana Bartlett 4-17-19
Approved by: Quality Manager or Designee	Gabriella Jaynes	[Signature] 4-17-19